

Babytherm[®] 8000 WB

Warming bed

Instructions for Use



Table of contents

For Your Safety and that of Your Patients	3	Troubleshooting	31
Safety instructions	3	Mattress heating	31
Intended Use	4	Height adjustment (optional)	31
Preparation	5	What's what	32
Trolley	5	Complete unit	32
Storage options	5	Mattress heater control panel	33
Bronchial aspirator	6	Technical Data	34
Infusion bottle holder	7	Construction and Description	36
Accessories for oxygen therapy	7	Heated mattress	36
Warming bed	10	Order List	37
Fitting the inner walls	10	Parts list	38
Fitting the ventilation hose bracket	11	Index	39
Mattress	11		
Bed canopy	11		
Tilting the bed	12		
In-hospital transport	12		
Testing Readiness for Operation	13		
Before using for the first time	13		
Before each use	13		
Operation	14		
Patient care	14		
Heat therapy	14		
Mattress heating	15		
Radiant heater	19		
Recommended heater settings	20		
Using the bed canopy	23		
Oxygen therapy	25		
Nebuliser therapy	27		
Phototherapy	27		
Bronchial aspiration	28		
Taking X-rays	28		
Care	29		
Dismantling	29		
Disinfecting/Cleaning/Sterilising	29		
Before next use	29		
Maintenance intervals	30		

For Your Safety and that of Your Patients

Strictly follow the Instructions for Use

Any use of the apparatus requires full understanding and strict observation of these instructions. The apparatus is only to be used for purposes specified here.

Maintenance

The apparatus must be inspected and serviced regularly by trained service personnel at six monthly intervals.

Repair and general overhaul of the apparatus may only be carried out by trained service personnel.

We recommend that a service contract be obtained with DrägerService and that all repairs also be carried out by them.

Only authentic Dräger spare parts may be used for maintenance.

Observe chapter "Maintenance Intervals".

Accessories

Do not use accessory parts other than those in the order list.

Liability for proper function or damage

The liability for the proper function of the apparatus is irrevocably transferred to the owner or operator to the extent that the apparatus is serviced or repaired by personnel not employed or authorized by DrägerService or if the apparatus is used in a manner not conforming to its intended use. Dräger cannot be held responsible for damage caused by non-compliance with the recommendations given above. The warranty and liability provisions of the terms of sale and delivery of Dräger are likewise not modified by the recommendations given above.

Dräger Medical AG & Co. KGaA

Safety instructions

Not for use in areas of explosion hazard

This apparatus is neither approved nor certified for use in areas where combustible or explosive gas mixtures are likely to occur.

Safe connection with other electrical equipment

Electrical connections to equipment which is not listed in these Instructions for Use should only be made following consultations with the respective manufacturers or an expert.

Do not use mobile phones within a distance of 10 metres from the machine.

Mobile phones can cause interference to electrical and electronic medical appliances, thereby putting patients at risk.*

* Dräger medical appliances comply with the interference immunity requirements of the specific standards for the products or EN 60601-1-2 (IEC 601-1-2). However, depending on the design of the mobile phone and situation of use, field strengths may occur in the immediate environment of a mobile phone that exceed the values of the standards quoted and therefore cause interference.

Intended Use

Babytherm® 8000 WB* is a unit for warming premature babies, neonates and infants with a body weight of up to 5 kg.

The unit may be used in operating theatres, neonatal wards, children's wards and paediatric intensive care units.

The unit may be operated by clinical staff or by nursing staff under the supervision of a doctor.

Babytherm 8000 WB provides heat through a heated gel mattress.

Therapy and nursing uses

- Warming up, maintaining warmth and thermal stability with mattress heating
- Cooling patients with high temperatures (temperature of mattress below body temperature)
- Adjustable bed angle for head-up or head-down position
- Transitional bed for weaning infants from incubators
- Oxygen and nebuliser therapy (with appropriate accessories)

Babytherm 8000 WB monitors and adjusts the temperature of the heating plate to attain and maintain a preselected temperature at the contact surface between the infant and the mattress.

The heat supplied through the mattress is not sufficient on its own for unclothed/uncovered infants/premature babies. An additional radiant heater or a Babytherm 8000 OC should be used.

Options:

- Bed canopy (for oxygen and nebuliser therapy)
- Height adjustment
- Swivel cabinet

The basic unit can be combined with the accessories shown on the Order List.

Optional therapy accessories:

- Secretion aspirator
- O2 flowmeter
- O2 monitor
- O2 hood
- O2 injector
- Infusion bottle holder
- Phototherapy unit
- Radiant heater on extra mobile stand

* Babytherm® 8000 WB – is a registered trademark of Dräger

Preparation

Trolley


Adjusting the height

Height adjustment is optional.

Before each use check:

- Connect power supply.
- 1 Press right pedal briefly to raise the bed.
- 2 Press left pedal briefly to lower the bed.

- Adjust to a comfortable working height.
- Do not place any objects in the way of height adjustment.



Storage options

Units with height-adjustable column:

- Swivel cabinet with 2 swivel compartments and 1 swivel tray (optional).

Units without height adjustment:

- 2 open compartments in the column,
- 1 swivel cabinet with 2 swivel compartments and 1 swivel tray (optional).

Bronchial aspirator

- The aspirator should be prepared and tested according to the specific instructions for use.

Positioning the ejector on the standard rail

- 1 Clip the ejector to the standard rail.

- Mounting bracket with projecting support:

For units with height adjustment:

- 2 Screw bracket to the column so that the bottle holder is outside the path of the swivel cabinet.

Units with a fixed column:

- Screw bracket to left side of column.


- 3 Place the bottle holder on the projecting support.

- 4 Connect the hose between the secretion bottle and the ejector.

- Clip the hose bracket to the standard rail.
- Mount the aspiration hose and attach to hose bracket.
- Screw on the connecting hose and plug the connector into the socket of the central supply system (park position).

When using the NIST* connecting hose:

- Call in an expert to fit the fixed adapter to the device connector.



Mounting the ejector to the bed

- Mounting bracket with projecting support:

For units with height adjustment

- 5 Screw the bracket to the column so that the bottle holder is outside the path of the swivel cabinet.

Units with a fixed column


- Connect bracket to left side of column.

- 6 Place the bottle holder on the projecting support.

- Connect the hose between the secretion bottle and the ejector.
- Mount the aspiration hose and attach to hose bracket.
- Screw on the connecting hose and plug the connector into the socket of the central supply system (park position).

When using the NIST* connecting hose:


- Call in an expert to fit the fixed adapter to the device connector.



* NIST: Non Interchangeable Screw Thread

Infusion bottle holder


- Hang holder from standard rail and secure in place.



Accessories for oxygen therapy

O₂ flowmeter


- Preparing the flowmeter:
 - 1 Check that the ball moves freely in the metering tube.
 - 2 Check that the sealing ring in the union nut is present and undamaged.
 - 3 Close valve clockwise.



- To mount the flowmeter on the standard rail:
 - 4 Press the slider all the way down; hook the flowmeter to the rail and release the slider.
 - 5 Screw on the O₂ connecting hose.
 - 6 Insert the connector of the connecting hose into the O₂ delivery socket and press it all the way in.

When using the NIST* connecting hose:

- Call in an expert to fit the fixed adapter to the device connector.




* NIST: Non Interchangeable Screw Thread


Nebuliser

Installing the nebuliser:

- 1 Check that the sealing ring above the nozzle is present and undamaged.
- 2 Insert atomiser on to housing as far as it will go.
- 3 Slide hook ring on to nebuliser bottle.
- Fill nebuliser bottle with aqua dest. up to the mark (quantity 220 mL).
- 4 Screw nebuliser bottle to housing.




- 5 Screw the nebuliser securely to the flowmeter.
- Fit the spiral hose to the nebuliser port.



O₂ supply via injector


- 1 Press locking lever on mounting flange and push injector into flange.
- Allow locking lever on injector to engage. The injector is now locked.
- Connect up hose.
- Fit bed canopy (see page 23).

O₂ supply should only be used with O₂ concentration monitoring, e.g. via Dräger Oxydig.

**O₂ monitor**

e.g. Dräger Oxydig

- Prepare the O₂ monitor in conformity with its specific Instructions for Use.
- Fasten O₂ monitor, complete with holder, to the standard rail.
Turn the O₂ monitor so that the control panel of the measuring unit faces the mattress heater control panel.
- Place O₂ sensor in the warming bed and feed the cable through the hole in the end panel.
- Plug the connector of the sensor into the measuring unit.



Warming bed

The bed is enclosed by two side panels and two end panels that can be opened independently of one another. The end panels have holes and flexible grommets for hoses and cables.

Fitting the panels

- 1 Insert the lower mounting pins of the panel hinge-pieces into the guide slots on either side.
- Position the panel semi-upright.
- 2 Press the panel down firmly until the pins lock into position at the bottom of the slots.
- Fold the panel up into the vertical position and allow it to drop into the locking position.
- Fit all four panels in this way.


Make sure that the panels are seated correctly.

To open:

- Lift the panel all the way up until the upper pins come out of their slots, then fold the panel down.

With lively infants or older children:

- Use 230 mm high inner panels.




012 28 517

Fitting the inner walls

With lively children, the optional inner walls should also be used even when the side panels are closed.

- Insert the pins into the tapered holes and press down slightly until the inner panel is firmly in place.


Make sure the panels are seated correctly.



013 28 517

Fitting the ventilation hose bracket

- Insert the hose bracket into the hole in the left-hand end panel and tighten with knurled nut.



01142 8 517

Mattress


- When operating **with** mattress heater, the gel mattress (2M 20 827) should be used.
- Place gel mattress in the cot and cover it with a cotton sheet.
Prevent contact with sharp objects as these could damage the gel mattress.
Do not fold or kink the gel mattress. To transport the gel mattress, roll it up.
Always observe the gel mattress warming-up time.
Wait for about 1 hour to allow the gel mattress to warm up before placing a patient in the Babytherm.

Bed canopy

Optional

Higher side panels (230 mm) are recommended when using the bed canopy.


- Check that the bed canopy is closed.



015 28 517

When the bed canopy is not required:


- Attach bracket to standard rail.
- 1 Hang bed canopy from bracket.




016 28 517

Tilting the bed

- 1 Pull the handle of the locking mechanism out towards the front.



- 2 Press handle down = head-up position
Pull handle up = head-down position.
- Release the handle: the bed automatically locks into the selected position.



Maximum tilt angle:

Head-up position:	20°
Head-down position:	15°

The bed can be tilted in finely graduated steps.

Preferred positions:

- Horizontal,
- 10° tilt head-up and
- 10° tilt head-down.

The mattress may slip at maximum tilt.

- In this case do not place a sheet over the mattress.

In-hospital transport

- If optional height adjustment is fitted, lower the unit to its lowest position.
- Swivel the swivel compartments or tray inwards.
- Accessories projecting beyond either side of the unit must be removed or folded in.
- Switch off heating systems and disconnect the mains plug.
- Disconnect gas supply hoses from central supply unit.
- Fit optional bed canopy to protect patient from draughts.

Immediately after transport:

- Plug in the mains plug and switch on the heating systems.

Testing Readiness for Operation

Before using for the first time

- Check that the mains voltage matches the values indicated on the rating plate next to the power cable.

Before each use

- Check that the unit has been disinfected.
- Check that the side panels are locked securely into place.
- Check that the side panels are free from cracks and sharp edges.
- Check that the correct mattress is in the cot;
Operation with mattress heater: gel mattress.
- Check that the bed tilts properly and locks securely into position.
- Check that the required accessories and therapy equipment are available and in proper working order.
- Check that the gas supply is available and sufficient for the accessories and equipment to be used.
- Check that the cables and hoses are correctly and securely installed.

Never route cables or hoses over the panels because they might be pinched or crushed when folding up the panels or fitting the bed canopy.

- Check that the therapy equipment used is ready for operation:

See chapter on heat therapy, see page 14

See chapter on oxygen therapy, see page 25

See chapter on phototherapy, see page 27

Operation

Patient care

Precautions

- Never leave the patient unattended when the side panels are down. **Risk of infant falling out of the cot.**
- When folding up the side panels, take care not to pinch any parts of the patient's body or any hoses or other articles, e.g. bedding.
- Check that all 4 panels are closed and securely locked in position: the yellow panel hardware should be visible above the surface of the bed.
- The inner walls should be used with all patients, whether the side panels are open or closed.
- With lively infants and/or older children, the 230 mm high side panels should be used. Such patients should be kept under very close observation.

Heat therapy

Two heating systems are available for keeping patients warm and if necessary for warming up or cooling down patients:

- Mattress heater for operation, see page 16.
- Radiant heater with extra mobile stand (accessory).
Following the specific instructions for use.

Precautions

- **Constantly monitor the core temperature of the patient.**
- **Adjust the temperature settings to the needs of the patient.**
- Increased heat is directed to the patient when the mattress heater and radiant heater are operated in combined mode. Follow the instructions on page 20 onwards for setting the two heating systems, separately and in combination.
- **When mattress heating is in use, the gel mattress must be in place on the heated bed area.**
- **The gel mattress should only be used as a heating pad with the Dräger Babytherm 8000. Do not use it with the Dräger Babytherm 4200 or other warming beds with heating plate temperatures over 40 °C – danger of burns.**
- Prevent contact with sharp objects, as these can damage the gel mattress. Do not fold or kink the gel mattress.
To transport the gel mattress, roll it up.
- **Always adhere to the gel mattress warming-up time.**
Wait for about 1 hour before placing a patient in the Babytherm, to allow the gel mattress to warm up sufficiently.
- Additional heat sources, such as phototherapy lamps or radiant heaters, increase the heat directed to the patient and should be taken into account. Avoid exposure to direct sunlight.

Mattress heating


Checking readiness for use



Before each use


- Check that the gel mattress has been disinfected and is in perfect condition.
- Lay gel mattress flat on the aluminium heating plate.
- Connect mains power.

Switching on and running self-test


- 1 Press main switch until it engages = ON.
The microprocessor checks the function of the device as part of the self-test.
- All the LEDs light up; the digital display reads **88.8** and an audible alarm sounds.
- 2 After approximately 2 seconds the device flashes to propose a temperature setting of 37.0 °C.




- 3 The display alternates cyclically between the letters **SEt** and the measured temperature value.
- 4 Press  or  key briefly to accept the temperature setting and
- 3 the display continuously shows the measured value = measured mattress temperature value.




- 5 If **Acc** is displayed alternately with the measured temperature value and at the same time an intermittent alarm sounds, this indicates that **the power failure alarm battery is insufficiently charged**.
After approximately 30 minutes operation, the battery is charged up. The display continuously shows the measured temperature value and the intermittent alarm is switched off.




To mute the intermittent alarm:

- 6 Press  key. The yellow LED lights up.
The intermittent alarm can be muted for 15 minutes.

- 7 Display **Err** alternating with the measured temperature and
- 8 red LED **Inop** lit – **device malfunction**, see "Troubleshooting" on page 31.




Testing the LEDs, displays and audible alarm:

- 1 Press  key:
for approximately 2 seconds – all the LEDs light;
the digital displays read **88.8** and an audible alarm sounds.


After that the displays and LEDs are dark and the audible alarm is switched off. After approximately 2 seconds the temperature is displayed again.


The function test can also be run while the device is operating.



- Check at least once a day.

Testing the power failure alarm

- Disconnect mains plug.
- 2 Red LED  lights, continuous alarm sounds.
- Reconnect mains plug.
The device carries on operating with the values set **before** the power failure alarm.



Operation

- Heat the mattress to 37 °C. Allow a warming-up time of around 1 hour.
During the warming-up stage and when in standby mode, the gel mattress should be covered with a blanket.
- Cover the gel mattress with a sheet before placing a patient on it.

If the radiant heater is not in use:



- Cover the patient with a blanket.


Do not switch off the mattress heater while an infant is lying on the gel mattress. Risk of hypothermia.

Setting the temperature

The set temperature can be adjusted in increments of 0,1 °C.

To set a temperature within **the normal range of 35 °C to 37 °C**:

- 1 Hold down  or  key until the desired setting is displayed.








To set a temperature **outside** the normal range:

- Extended lower limit: **30 °C to 35 °C**,
- Extended upper limit: **37 °C to 38.5 °C**.

Only use low temperatures if prescribed by doctor.


Monitor patient very closely.

- 2 Press unlocking key , yellow LED in the key lights up.
 - 3 Hold down  or  key until the desired setting is displayed.
- The setting over 37 °C or under 35 °C should be set within 1 minute. Otherwise the LED in the key  is extinguished. The extended setting range is locked again.



If the temperature is outside the measuring range of 5 °C to 45 °C:

- 3 dashes at the **bottom** of the display = temperature below 5 °C
 - 3 dashes at the **top** of the display = temperature above 45 °C.
- See chapter on "Troubleshooting", page 31.




Alarms


Deviation from set temperature

If the deviation between the set and measured mattress temperature is greater than ± 1 °C:

- 1 The red LED ± 1 °C starts flashing
- an intermittent alarm sounds and
- 2 the display flashes.




The intermittent alarm can be muted for 15 minutes:

- 3 Press  key. The yellow LED in the key lights up
- 1 the red LED is lit continuously.

When the measured mattress temperature returns to within ± 1 °C of the set value:

- The red LED ± 1 °C goes out and the intermittent alarm is switched off.

After initially switching on the mattress heater:


- the audible alarm is suppressed for one hour (warming-up phase).
- 1 Red LED ± 1 °C lights up.
 - 3 Yellow LED in the  key lights up.

40 °C alarm

- 4 Red LED 40 °C flashes and a continuous alarm sounds.
- 5 The display flashes.

The continuous alarm cannot be suppressed.

The alarm stops automatically as soon as the mattress temperature drops to the normal setting.




Shutting down

Only for care purposes, when changing over patients.

- Press main power switch all the way home = OFF then
 - Disconnect mains plug.
- It is important to proceed in this order, as otherwise the power failure alarm is activated.

Radiant heater

- Slide the radiant heater mobile stand underneath the Babytherm 8000 WB.
- Follow the instructions for use of the radiant heater. See Page 20 for settings.





Recommended heater settings

Instructions for using and setting the mattress heater and radiant heater, alone or in combination.


Using the mattress heater with a gel mattress without using the radiant heater

Always monitor the core temperature of the patient and adjust the mattress temperature setting to the needs of the patient.


- Always cover and/or clothe the patient.
Do not place blankets or other insulating material under the patient, since they obstruct heat transfer to and from the bed (warming/cooling).




- To warm the patient:
Set the temperature to e.g. 38.5 °C.



- To neither warm nor cool the patient:
Set the temperature to 36 °C to 37 °C.




- To cool the patient:
Set the mattress temperature to e.g. 36 °C or lower.




Using radiant heater only

Always monitor the core temperature and adjust the temperature setting to the needs of the patient.

- Observe the minimum distance (70 cm) between the underside of the radiant heater and the bed surface.
- The radiant heater is less effective when the bed canopy is used.
(The acrylic glass bed canopy provides only limited permeability to the infrared radiation of the radiant heater).




- Do not cover or clothe the patient.




Using the radiant heater in combination with mattress heater and gel mattress


- To provide heat from above in addition to the mattress heater.
Set the mattress heater temperature e.g. to 38 °C and set the skin temperature setting on the radiant heater e.g. to 36.3 °C.



- No heat transfer to or from the patient from below:
Set the mattress heater temperature to 36 °C to 37 °C and set the skin temperature setting on the radiant heater e.g. to 36.3 °C.



- To cool the patient from below:
Set the mattress heater temperature e.g. to 36 °C or lower and set the skin temperature setting on the radiant heater e.g. to 36.3 °C.



Using the bed canopy

When the bed canopy is in place, the side panels can be folded down. **Do not leave the Babytherm unattended with the side panels open, as the infant can fall out.**

It is advisable to use the higher side panels (230 mm) when using the bed canopy.


When using external phototherapy, remember the maximum load on the bed canopy is 11 kg.

The bed canopy should **not** be used as a convenient place to lay objects, clothing, etc.


Installing the bed canopy

Check whether the bed canopy is in the closed position; otherwise:

- 1 Close bed canopy = turn lock until it engages.




- 2 Hold bed canopy handles with both hands and place over the side panels.




Opening/closing the bed canopy

- 1 Unlock bed canopy = turn lock.
- 2 Open bed canopy = swivel handles upwards until they engage. The infant is now accessible for care/medical treatment.
- 2 Close bed canopy = swivel handles forwards.




Removing the bed canopy

- 3 Put bed canopy in closed position = turn lock until it engages.
- Remove bed canopy by holding the handles with both hands (follow instructions on the bed canopy).



- 4 Hang bed canopy from the bracket.



Oxygen therapy

Precautions

The risks of the system are increased when oxygen is in use.

- Avoid naked flame and lit cigarettes. Textiles, plastics and oils are more easily ignited and burn with greater intensity in an oxygen-rich atmosphere.
- Keep oxygen fittings and seals free of oil and grease.
- Open the valves slowly.
- Do not use Babytherm 8000 WB in the presence of flammable anaesthetic gases or disinfectants.
- Do not use or store flammable liquids such as alcohol, ether or acetone in the Babytherm 8000 WB.
- Do not use electrical appliances underneath the Babytherm bed canopy, except for devices authorised and certified for operation in explosion-hazard areas.

Physiological dangers of oxygen

Only enrich the oxygen concentration under the instructions of a doctor and only in accordance with the arterially measured oxygen partial pressure in the patient's blood. Otherwise there is a danger of hyperoxaemia (eye damage) or hypoxaemia (brain damage).

During oxygen therapy, the oxygen concentration must be constantly monitored, e.g. with Dräger Oxydig.

Increased noise level

Patients may suffer excessive noise when compressed gas is administered using a hood.

Administering oxygen using the injector

(in the whole interior space)

- Place canopy in position (see page 23).
- 1 Set O₂ concentration on the injector:
30, 40, 50 vol.% O₂.
- Set O₂ supply on flowmeter.


Recommended settings:

Injector setting	Vol.% O ₂	30	40	50
O ₂ supply	L/min	5	10	15

- 2 Monitor the O₂ concentration, e.g. with Dräger Oxydig.

After use

- Open the bed canopy.
- Close the flow valve on the O₂ flowmeter = turn clockwise.
- Dismantle and service the oxygen therapy equipment, see page 29.




041 26 512

Administering oxygen using the hood

- 3 Place the Oxydig sensor with the adapter ring in the star-shaped cuff cut-out.
Plug the connector into the measuring device.
- 4 Connect the hose to the hood.
- Lay the patient on his/her back.
- 5 Place the hood over the patient's head.
The lower edge must rest on the mattress.
Do not seal the collar area.
- 6 Set the O₂ supply at the metering valve: 0.5 to 4 L/min.
- 7 Monitor the O₂ concentration, e.g. with Dräger Oxydig.

After use

- Remove the hood from the patient.
- Close the flow valve on the O₂ flowmeter = turn clockwise.
- Dismantle and service the oxygen therapy equipment, see page 29.




043 26 517

Nebuliser therapy

Using the McGaw nebuliser.

Use only with the Babytherm 8000 WB closed.

- Place the canopy in position (see page 23).
- Prepare the McGaw nebuliser and heater according to its instructions for use and attach to standard rail of Babytherm 8000 WB.
- Lead the spiral hose through one of the flexible grommets in the end panel into the interior.
- Use the McGaw nebuliser according to its instructions for use.



Phototherapy

Use on a stand

- Slide the stand underneath the trolley of the Babytherm 8000 WB.
- Adjust height of phototherapy unit: distance between underside of lamp body and the patient must be **not less than 0.3 metres**.


When using a radiant heater at the same time:

- Slide the mobile radiant heater underneath the trolley of the Babytherm 8000 WB.
- Position the housing so that the patient receives the radiant heat directly.
- Follow the instructions for use of the phototherapy unit.

When using the radiant heater in the »Man« mode:

- Reduce the radiant heater output level by about 3 increments compared to operation without phototherapy.


The phototherapy light directs additional warmth to the patient.
Monitor the patient's core temperature regularly.



Use without a stand

- Place the 230 mm high side panels in position.
- 1 Place the canopy in position (see page 23).
- 2 Place Photo-Therapy 4000 on the canopy.
- Follow the instructions for use of the Photo-Therapy 4000.

Monitor the patient's core temperature regularly.



Bronchial aspiration

- Use bronchial aspirator in accordance with its specific instructions for use.

Taking X-rays

- Place the X-ray film cartridge directly underneath the patient.
- Position the X-ray machine above the Babytherm 8000 WB.

If a radiant heater is in use at the same time:

- Swivel the radiant heater to one side.
- After taking the X-ray, swivel the radiant heater back into position to avoid chilling the patient.

Care

The Babytherm 8000 WB infant warmer system must be thoroughly cleaned and disinfected

- after each change of patient
- at least once a week.

Clean and disinfect all accessories, e.g. bronchial aspirator, in accordance with their specific Instructions for Use.

Dismantling

- Switch off the device(s). Disconnect the power plug from the mains and switch off all compressed gas supplies used.
- Remove any ancillary equipment installed.
- Remove the gel mattress from the cot.
Always store the gel mattress flat.
- Remove the ventilation hose clips.
- Remove the silicone grommets.

Disinfecting/Cleaning/Sterilising

Bed frame, inside and outside;

Side panels, inside and outside;

Inside walls, inside and outside;

Bed surface;

Gel mattress;

Bed canopy;

Trolley, including all attachments:

- Wipe off visible soiling with a disposable cloth wrung out in a detergent.
- Disinfect surfaces by wiping with disinfectant.
- After waiting the prescribed time for the disinfectant to act, wipe with a clean damp cloth and dry.
- **Do not wash the mattress in a washing machine.**
Do not autoclave.

Silicone grommets:

- Wash with detergent and rinse with clean water.
- Disinfect the components in a disinfectant bath. After immersing for the prescribed time, rinse with clean water and dry; or
- Sterilise at 120 °C (glove programme).

Gel mattress:

- We recommend wipe disinfecting the gel mattress with Incidin Extra (Henkel, Düsseldorf)

Do not expose Babytherm 8000 WB to UV radiation. Cracks may be caused in the acrylic glass components.

Only use the recommended cleaning agents and disinfectants. Otherwise, there is a danger of causing cracks in the acrylic glass and macrolon, e.g. if alcohol is used.

Use only preparations classified as "surface disinfectants" for disinfecting. For material compatibility, we recommend preparations based on

- aldehydes,
- quaternary ammonium compounds.

Due to the risk of damage to materials, the following preparations are unsuitable:

- halogen-releasing compounds,
- strong organic acids,
- oxygen-releasing compounds.

The manufacturer's recommendation should always be followed when choosing the product. The manufacturer is liable for the data concerning suitable applications for the products and any damage to property.

Do not use disinfectants or cleaning products that contain alcohol.

For users in the Federal Republic of Germany, we recommend the use of disinfectants listed in the current DGHM list (DGHM = German Society for Hygiene and Microbiology). The DGHM list (published by mhp-Verlag GmbH, Wiesbaden) also specifies the active basis of each disinfectant. For countries where the DGHM list is unavailable, the above recommendations apply.

For example, the following wipe-disinfectants may be used:

Bacillotox	Bode, Hamburg
Incidur	Henkel Hygiene GmbH, Düsseldorf
Sekusept-Pulver	Henkel Hygiene GmbH, Düsseldorf

Follow the Instructions for Use provided by the disinfectant manufacturer.

Before next use


- Reassemble components.
- After wipe-disinfecting, operate the fully assembled unit for a few hours without a patient to eliminate any disinfectant residues.
(Mattress heater temperature setting = 37 °C)

Before a patient is next placed in the unit:

- Fit all therapy accessories required, see page 6.
- Check that the unit is ready for operation, see page 13.

Maintenance intervals

Always disinfect and clean the unit and accessories before any maintenance * – even when returning the unit to the supplier for repair.



 Always disconnect power supply before any maintenance!

Gel mattress	Patch over small cuts or other damage in the covering film with adhesive tape. Use the gel mattress repair kit. Replace the gel mattress if the material becomes brittle or large cracks appear; replace after 2 years at the latest.
Silicone grommets	Replace as soon as the material becomes brittle and/or sticky.
Inspection and maintenance	Yearly by trained service personnel.

* Definitions:
Inspection = examination of actual condition
Service = measures to maintain specified condition
Repair = measures to restore specified condition
Maintenance = inspection, service, repair

Troubleshooting

Mattress heating


Message	Cause	Remedy
Red LED ± 1 °C flashing, measured value display flashing, intermittent alarm on.	Measured value deviates from set value by more than ± 1 °C.	When using in combination with radiant heater: Reduce the output level of the radiant heater.
Red LED 40 °C flashing, measured value display flashing, continuous alarm on.	Mattress temperature above 40 °C	When using in combination with radiant heater: Reduce the output level of the radiant heater.
Red LED  lit, continuous alarm on.	Power failure	Check that unit is plugged into mains. Check that mains is switched on. Call DrägerService.
Red LED Inop lit. The measured value display shows Err alternately with the measured value. The error code is flashing in the setting display. Continuous alarm sounds.	Malfunction	Call DrägerService.
The three middle segments of the measured value display are flashing.	Temperature sensors defective.	Call DrägerService.
The measured value display shows Acc alternately with the measured value. Intermittent alarm on.	Power failure alarm battery is insufficiently charged.	Operate unit for 30 minutes, press  key to mute alarm for 15 minutes.
Measured value display shows: Three dashes at bottom	Mattress temperature is under 5 °C	Wait until mattress temperature is above 5 °C
Three dashes at top	Mattress temperature is above 45 °C	Wait until mattress temperature is under 45 °C

Height adjustment (optional)

Message	Cause	Remedy
Height adjustment inoperative.	Motor overheat switch has tripped.	Allow to cool for about 15 minutes.


What's what

Complete unit



- | | |
|-----------------------------------------------------|-----------------------------------------------------|
| 1 Bed canopy | 9 Foot controls for electric height adjustment only |
| 2 End panel | 10 Chassis with 4 castors, lockable |
| 3 Side panel | 11 Accessory rail |
| 4 Handle with inner release bar for tilting the bed | 12 Inner wall, set of 2 |
| 5 Mattress heater control panel | 13 Silicone grommets |
| 6 Bed assembly | 14 O ₂ injector holder |
| 7 Swivel cabinet, optional | 15 Gel mattress with mattress cover |
| 8 Column, optional with electric height adjustment | |

Mattress heater control panel



- 1 Pushbutton for extending the setting range;
downwards: 30 °C to 35 °C,
upwards: 37 °C to 38.5 °C
- 2 Display of set (desired) mattress temperature;
on the left: key for reducing the setting,
on the right: key for increasing the setting.
- 3 Pushbutton for muting the intermittent alarm for 15 minutes;
the yellow LED remains lit while the alarm is muted.
- 4 Pushbutton for testing LEDs, displays and audible alarm.
- 5 On/off switch
- 6 Measured mattress temperature display.
- 7 Red LED ⏻ ;
lit following a power failure.
- 8 Red LED 40 °C;
flashes if the measured value is above 40 °C.
- 9 Red LED ± 1 °C;
flashes if the measured mattress temperature deviates from
the set value by more than ± 1 °C.
- 10 Red LED Inop;
indicates device malfunction.

Technical Data

Ambient conditions

For operation	
Temperature	20 °C to 30 °C
Atmospheric pressure	900 hPa to 1100 hPa
rel. humidity	0 to 90 %, no condensation
For storage/transport	
Temperature	−20 °C to 60 °C
Atmospheric pressure	900 hPa to 1100 hPa
rel. humidity	0 to 90 %, no condensation

Mattress heater


Temperature measurement	
Measuring range	5 °C to 45 °C
Accuracy	±0.5 °C between 20 and 42 °C ±2 °C in all other parts of range
Setting range	30 °C to 38.5 °C

Height adjustment (optional)

Height adjustment range	295 mm
Specified operation	Intermittent operation 1 min. operation followed by 60 min. cooling period

Operating data

Mains voltage	100/120/127/230/240 V (to be specified on order) 50/60 Hz
Power consumption	
Mattress heater	160 W
Height adjustment (optional)	560 W
Fuses	
Mattress heater	
Range 100 V to 127 V	T 1,6 A L 250 V; IEC 127-2/III (2 units)
Range 230 V to 240 V	T 800 mA L 250 V; IEC 127-2/III (2 units)
Height adjustment (optional)	
Range 100 V to 127 V	T 6,3 A L 250 V; IEC 127-2/III (2 units)
Range 230 V to 240 V	T 3,15 A L 250 V; IEC 127-2/III (2 units)

Electrical protection class	I
Heating surface of mattress	 Typ BF
Standards	EN 60601-1 62 D (Sec) 146 IEC 601-2-35 Part 2, for mattress heater
Classification (according to EC Directive 93/42/EC Appendix IX)	Class IIb
UMDNS-Code (Universal Medical Device Nomenclature System)	15-610
Electromagnetic compatibility (EMC)	according to EN 60601-1-2
Dimensions	
Length x width	945 x 620 mm
Overall height (with bed canopy)	
Unit without height adjustment	1295 mm
Unit with height adjustment	1230 to 1515 mm
Working height of bed surface	
without height adjustment	950 mm
with height adjustment	885 to 1180 mm
Bed dimensions	750 x 490 mm
Bed tilting	adjustable in small increments maximum of 20° front end down maximum of 15° front end up
Height of side panels	230 mm or 150 mm
Height of inner walls	70 mm
Weight (with cabinet and bed canopy)	
Unit without height adjustment	85 kg
Unit with height adjustment	83 kg

Construction and Description

Heated mattress

The heated mattress consists of the following components: gel mattress, aluminium heating plate and an electronic monitoring and control unit.

The gel mattress consists of a highly heat-conductive gel that does not run or dislocate even when the bed is tilted.

The gel is surrounded by a film of soft material compatible with the skin.

The gel mattress moulds itself to the body contours of the patient lying on it, thereby providing a large contact surface that transfers warmth to the patient and avoids pressure points (bedsores).

When the gel mattress is cold, warmth is transferred from the patient to the mattress. The gel has good heat storage properties. If the mains supply is interrupted, e.g. for in-hospital transport, the patient will be kept warm for about 15 minutes, provided the insulation is sufficient (blanket, bed canopy).

The heating system consists of a 6 mm thick aluminium plate and a heater element below the plate. This system ensures that heat is distributed evenly over the entire bed surface.

The control and monitoring unit is located below the heating system. The touch-sensitive keypad with the pushbuttons and displays is arranged for convenience of operation.

The mattress temperature is adjusted with reference to a set temperature entered by the user. The set temperature is only attained in the areas where the patient is in contact with the mattress.

System monitoring includes self-check routines and alarms.

Order List

Name and Description	Part No.	Name and Description	Part No.
Babytherm 8000 WB, 230V (with inner walls)	FR 00 050	Therapy accessories	
Babytherm 8000 WB, 120V (with inner walls)	FR 00 088	Bronchial aspiration	
Mains voltage to be specified with order		Bronchial aspiration unit, complete, ejector up to 0.5 bar, with rail clamp	2M 85 045
Variants to be specified with order		Bronchial aspiration unit, complete, ejector up to 0.9 bar, with rail clamp	2M 85 040
— Set of side panels, height 230 mm or Set of side panels, height 150 mm		Bronchial aspiration unit, complete, vacuum up to 0.9 bar, with rail clamp	2M 85 041
— Mattress heater with gel mattress		Hose bracket	2M 85 239
— Stand with electric height adjustment or stand with fixed column		Bronchial aspiration unit, ejector up to 0.5 bar, with assembly kit for installation under the bed	2M 21 187
Options:		Bronchial aspiration unit, ejector up to 0.9 bar, with assembly kit for installation under the bed	2M 21 188
— Swivel cabinet		Bronchial aspiration unit, vacuum up to 0.9 bar, with assembly kit for installation under the bed	2M 21 189
— Bed canopy			
Options (for retrofitting)		DIN connecting hoses:	
Set of side panels, height 230 mm	2M 21 034	O ₂ /air connecting hose 1.5 m	M 29 283
Set of side panels, height 150 mm	2M 21 032	O ₂ /air connecting hose 3 m	M 29 243
Inner panel	2M 20 936	O ₂ /air connecting hose 5 m	M 29 263
Swivel cabinet	2M 20 638	Air connecting hose 5 m	M 29 259
Cabinet fixing	2M 20 868	Vacuum connecting hose 1.5 m	M 29 287
		Vacuum connecting hose 3 m	M 29 247
		Vacuum connecting hose 5 m	M 29 267
Accessories		NIST connecting hoses, colour neutral:	
Infusion bottle holder	2M 16 520	O ₂ /air connecting hose 1.5 m	M 34 410
Inner rail, additional for drainage bag	2M 21 150	O ₂ /air connecting hose 3 m	M 34 411
Ventilation hose bracket	2M 21 191	O ₂ /air connecting hose 5 m	M 34 412
Shelf, 300 x 200 mm with rail clamp for standard rail	2M 24 678	Air connecting hose 5 m	M 34 409
Bed canopy	2M 30 392	Vacuum connecting hose 1.5 m	M 34 413
Bed canopy bracket	2M 21 342	Vacuum connecting hose 3 m	M 34 414
		For DIN/NIST conversion:	
		Fixed adapter O ₂ /AIR NIST/DIN	M 32 493
		Fixed adapter AIR NIST/DIN	M 32 495
		Fixed adapter vacuum NIST/DIN	M 32 496

Name and Description	Part No.
Phototherapy	
Photo-Therapy 4000, 230 V	2M 21 000
Photo-Therapie 4000, 120 V	2M 21 700
Phototherapy unit stand	2M 21 190
Heat therapy	
Radiant heater 600 W, on mobile stand	2M 20 539
Radiant heater 600 W, wall-mounted	2M 20 116
Wall mounting kit	2M 20 540
Oxygen therapy in interior space	
Injector; 30, 40, 50 vol.% O ₂ selectable	2M 14 190
O ₂ flowmeter, 0 to 16 L/min	2M 85 501
Nebuliser	2M 85 835
Spiral hose, 1 m (set of 5)	2M 17 478
Oxygen therapy with hood	
O ₂ hood	2M 19 250
O ₂ flowmeter, 0 to 16 L/min	2M 85 501
Nebuliser	2M 85 835
Spiral hose, 1 m (set of 5)	2M 17 478
Nebuliser therapy	
McGaw nebuliser, O ₂	2M 19 030
O ₂ hose 0.8 m, McGaw	2M 85 563
Nebuliser heater, McGaw	2M 18 850
Spiral hose, 1 m (set of 5)	2M 17 478
DIN connecting hoses:	
O ₂ connecting hose 1.5 m	M 29 271
O ₂ connecting hose 3 m	M 29 231
O ₂ connecting hose 5 m	M 29 251

Name and Description	Part No.
NIST connecting hoses, side neutral:	
O ₂ connecting hose 1.5 m	M 34 401
O ₂ connecting hose 3 m	M 34 402
O ₂ connecting hose 5 m	M 34 403
Consumables	
Gel mattress for use with mattress heater	2M 20 827
Gel mattress repair kit	2M 21 324
Mattress cover	2M 21 272
Silicone grommet	2M 20 434
For bronchial aspiration	
Aspiration hose 1.5 m	M 25 780
Secretion inspection window, set of 5	M 07 582
Bacterial filter CH 102, set of 5	67 23 976
Canister	M 20 091
Bottle cap with valve, plastic	2M 85 012

Parts list

Instead of the part nos. shown in the Order List, the following parts or units no longer in the Dräger product range can be used:

Name and Description	Part No.
Canopy	2M 21 030
Oxydig, O ₂ meter	83 04 411
Oxydig meter bracket	2M 17 770
O ₂ sensor capsule	68 50 645
Oxydig sensor housing	68 50 250

Index

Alarm	18	O2 monitor	9
Ambient conditions	34	Operation	14
Bed canopy		Options	4
Preparation	11	Order List	37
Using	23	Oxygen therapy	
Before each use	13	Precautions	25
Before next use	29	with hood	26
Bronchial aspirator	6, 28	with injector	26
		Oxygen therapy, prepare accessories	7
Care	29	Parts list	38
Cleaning	29	Patient care, precautions	14
Construction and Description	36	Phototherapy	27
		Precautions	14, 25
		Preparation	5
Disinfecting	29		
Dismantling	29	Radiant heater	19
		and mattress heater	22
Fitting the inner walls	10	Using	21
Fitting the panels	10	Recommended heater settings for heat therapy	20
For Your Safety and that of Your Patients	3		
		Safety instructions	3
Gel mattress	11	Setting the temperature	17
		Shutting down	18
Heat therapy	14	Sterilising	29
Recommended heater settings	20		
Infusion bottle holder	7	Table of contents	2
Injector	26	Technical Data	34
Intended Use	4	Testing Readiness for Operation	13
		Therapy and nursing uses	4
Liability	3	Tilting the bed	12
		Transport, in-hospital	12
		Trolley	5
		Troubleshooting	31
Maintenance	3		
Maintenance intervals	30	Using for the first time	13
Mattress heating	15		
Alarms	18	Ventilation hose, fitting bracket	11
and radiant heater	22		
Checking readiness for use	15	Warming bed, preparation	10
Operation	16	What's what	32
Mattress, prepare	11		
McGaw nebuliser	27	X-ray	28
Mobile phones	3		
Nebuliser therapy	27		

These Instructions for Use apply only to
Babytherm 8000 (FR 00 050)
with Serial No.:

If no Serial No. has been filled in by Dräger
these Instructions for Use are provided for
general information only and are not inten-
ded for use with any specific machine or
device.



Directive 93/42/EEC
concerning Medical Devices

Dräger Medical AG & Co. KGaA

Germany

🏠 Moislinger Allee 53 – 55
D-23542 Lübeck

☎ +49 451 8 82- 0

📠 26 80 70

FAX +49 451 8 82- 20 80

💻 <http://www.draeger.com>

90 28 521 - GA 6132.211 en

© Dräger Medical AG & Co. KGaA

4th edition - July 2001

Subject to alteration